

# Measurement of Moisture Content(M.A.)

**[SJ-MM-001]**

2005-05-25

Supersedes prior issues

**Quality Assurance**

## Test procedure

1. Completely dry a test dish for 3 ~ 4 minutes.
2. Take powder as a sample by 1 ~ 1.2g and put the sample on the dried dish.
3. Place the sampled dish into Moisture Analyzer for 130 °C, 30 minutes.
4. Read moisture content.

## Equipment

Sartorius MA30 (Moisture Analyzer)



**SUNJIN CHEMICAL CO.,LTD.**