

Measurement of Residual Solvent

[SJ-MM-031]

2014-08-04
REV.0

Quality Assurance

Test procedure

1. Stir it for 30 mins after adding sample 30g and DMF 90g into glass beaker
2. Measure G.C with the liquid after filtering the 1) (filtering paper : 125nm)