

전기전도도 측정

[SJ-MM-037]

2011-01-12
Supersedes prior issues

Quality Assurance

Test method

1. Stir for a minute after mixing sample 2g and distilled water 198g
2. Conduct the conductivity Meter(SJI-Q-020)
3. Read the result after 5 minutes

Equipment

Conductivity Meter (EC214)

